

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/08/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12020

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:44
INTERNAL STANDARD	VERIFIED	00:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:44
BLANK TEST	0.00	00:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:45
BLANK TEST	0.00	00:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:46
BLANK TEST	0.00	00:47

