

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/14/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12020

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:02
INTERNAL STANDARD	VERIFIED	01:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:03
BLANK TEST	0.00	01:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:04
BLANK TEST	0.00	01:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:05
BLANK TEST	0.00	01:06

