

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/23/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12020

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:25
INTERNAL STANDARD	VERIFIED	01:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:26
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:27
BLANK TEST	0.00	01:27
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:28
BLANK TEST	0.00	01:28



565 *Ralph Bartlett* WSO #162 05-23-12