

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 06/20/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:04
INTERNAL STANDARD	VERIFIED	01:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:05
BLANK TEST	0.00	01:05
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:06
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:07
BLANK TEST	0.00	01:08



SGT *Ralph Bartlett* WCSO SHERIFF #162 06-20-12