

SOLUTION CHANGE




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 07/03/2012

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:06
INTERNAL STANDARD	VERIFIED	02:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	02:07
BLANK TEST	0.00	02:08
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:08
BLANK TEST	0.00	02:09



*Sgt Ralph Bartlett*

WCSO Sheriff #162 07-03-12