

SUPERVISOR TEST




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 07/09/2012

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:39
INTERNAL STANDARD	VERIFIED	03:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:39
BLANK TEST	0.00	03:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:40
BLANK TEST	0.00	03:41
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:41
BLANK TEST	0.00	03:42



SG Bartlett

WCSO SHFAD #162 07-09-12