

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/26/2012

Operator Name: JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12060

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:56
INTERNAL STANDARD	VERIFIED	00:56
REFERENCE STANDARD TEMP	34.03C	
REFERENCE STANDARD	0.09	00:57
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	33.94C	
REFERENCE STANDARD	0.09	00:58
BLANK TEST	0.00	00:58
REFERENCE STANDARD TEMP	33.97C	
REFERENCE STANDARD	0.09	00:59
BLANK TEST	0.00	01:00

Joshua M Lopez
#299

