

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 07/31/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:03
INTERNAL STANDARD	VERIFIED	23:03
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:03
BLANK TEST	0.00	23:04
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:04
BLANK TEST	0.00	23:05
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:06
BLANK TEST	0.00	23:06



*SGT Rusty Bartlett WCSO #162 - 07-31-12*