

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/15/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:07
INTERNAL STANDARD	VERIFIED	01:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:07
BLANK TEST	0.00	01:08
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	01:08
BLANK TEST	0.00	01:09
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:09
BLANK TEST	0.00	01:10



505 Dept of Justice WCSO #162 08-15-12