

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 08/22/2012

Operator Name: JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:41
INTERNAL STANDARD	VERIFIED	18:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:41
BLANK TEST	0.00	18:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:42
BLANK TEST	0.00	18:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:43
BLANK TEST	0.00	18:44

