

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/11/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12080

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 00:43 |
| INTERNAL STANDARD       | VERIFIED | 00:43 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 00:44 |
| BLANK TEST              | 0.00     | 00:44 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.09     | 00:45 |
| BLANK TEST              | 0.00     | 00:45 |
| REFERENCE STANDARD TEMP | 33.97c   |       |
| REFERENCE STANDARD      | 0.09     | 00:46 |
| BLANK TEST              | 0.00     | 00:46 |



SGT *Ralph Bartlett* WCSO #102 09-11-12