

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/17/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:44
INTERNAL STANDARD	VERIFIED	04:45
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:45
BLANK TEST	0.00	04:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:46
BLANK TEST	0.00	04:47
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:47
BLANK TEST	0.00	04:48



SG5 Rugby Bunches WCSO #162 09-17-12