

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/24/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12080

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:00
INTERNAL STANDARD	VERIFIED	23:00
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:00
BLANK TEST	0.00	23:01
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	23:01
BLANK TEST	0.00	23:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:02
BLANK TEST	0.00	23:03



SGT Rust / Bartlett

WC50#162

09-24-12