

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/01/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:07
INTERNAL STANDARD	VERIFIED	01:07
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:08
BLANK TEST	0.00	01:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:09
BLANK TEST	0.00	01:10
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:10
BLANK TEST	0.00	01:11



SG- *Ralph Bartlett* WCSO #162 10-01-12