

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 10/09/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:36
INTERNAL STANDARD	VERIFIED	02:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:36
BLANK TEST	0.00	02:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:37
BLANK TEST	0.00	02:38
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:38
BLANK TEST	0.00	02:39



SGS *Ralph J Bartlett* WCSO #162 10-09-12