

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/23/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:59
INTERNAL STANDARD	VERIFIED	02:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:00
BLANK TEST	0.00	03:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:01
BLANK TEST	0.00	03:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:02
BLANK TEST	0.00	03:03



SGS Dept of Justice WSO #162 10-23-12