

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/29/2012

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN COUNTY SHERIFF'S OFFICE
Reference Standard #: 12110

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:36
INTERNAL STANDARD	VERIFIED	01:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:36
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:37
BLANK TEST	0.00	01:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:38
BLANK TEST	0.00	01:39



SG *Ralph Bartlett* WCSO #160 10-29-12