

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/06/2012

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:13
INTERNAL STANDARD	VERIFIED	09:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:14
BLANK TEST	0.00	09:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:15
BLANK TEST	0.00	09:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	09:16
BLANK TEST	0.00	09:16

Charles B. Geisler #276

