

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 12/26/2012

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:04
INTERNAL STANDARD	VERIFIED	23:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:04
BLANK TEST	0.00	23:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:05
BLANK TEST	0.00	23:06
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:07
BLANK TEST	0.00	23:07



SCS *Ralph J Bartlett*

WCSO SHERIFF #162 12-26-12