

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 01/07/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:25
INTERNAL STANDARD	VERIFIED	00:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:26
BLANK TEST	0.00	00:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:27
BLANK TEST	0.00	00:27
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:28
BLANK TEST	0.00	00:28



SGT Ralph J Bartlett wcsb #162 01-07-13