

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/15/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:30
INTERNAL STANDARD	VERIFIED	00:30
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:32
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:33



SGS *Ralph Bartlett* 01-15-13 WCSO #162