

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/21/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:44
INTERNAL STANDARD	VERIFIED	02:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:45
BLANK TEST	0.00	02:45
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:46
BLANK TEST	0.00	02:46
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:47
BLANK TEST	0.00	02:47



SGS Rpt of Bartlett

WCSO SHERIFF #162 01-21-13