

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/30/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:58
INTERNAL STANDARD	VERIFIED	00:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:59
BLANK TEST	0.00	00:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:00
BLANK TEST	0.00	01:00
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:01
BLANK TEST	0.00	01:01



SGR *[Signature]* WCSO #142 01-30-13