

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/07/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:17
INTERNAL STANDARD	VERIFIED	00:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:18
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:19
BLANK TEST	0.00	00:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:20
BLANK TEST	0.00	00:20



SGT Ralph Bartlett

WCSO SHELLO #162 02-07-13