

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/11/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:53
INTERNAL STANDARD	VERIFIED	23:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:53
BLANK TEST	0.00	23:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:54
BLANK TEST	0.00	23:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:55
BLANK TEST	0.00	23:56



565 Peak of Benzene was #42 02-11-13