

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 02/18/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:33
INTERNAL STANDARD	VERIFIED	00:33
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:34
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	00:34
BLANK TEST	0.00	00:35
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:35
BLANK TEST	0.00	00:36



SG- Ralph J Bartlett WCSO #162 02-18-13