

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/26/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:10
INTERNAL STANDARD	VERIFIED	00:10
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:10
BLANK TEST	0.00	00:11
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:11
BLANK TEST	0.00	00:12
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:12
BLANK TEST	0.00	00:13



SGS Rly Rwy Weso #162 02-26-13