

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 03/04/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:30
INTERNAL STANDARD	VERIFIED	01:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:30
BLANK TEST	0.00	01:31
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:31
BLANK TEST	0.00	01:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:32
BLANK TEST	0.00	01:33



SOT *Ralph J Bartlett*

WRSO #162

03-07-13