

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/17/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:02
INTERNAL STANDARD	VERIFIED	14:02
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	14:03
BLANK TEST	0.00	14:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:04
BLANK TEST	0.00	14:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:05
BLANK TEST	0.00	14:05



SGT *Ralph Bartlett* WCSO #162 03-17-13