

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/26/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 12230

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:41
INTERNAL STANDARD	VERIFIED	00:41
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:42
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:43
BLANK TEST	0.00	00:44



SG- Ryl / Bunter WISO #162 03-26-13