

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/01/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13050

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:06
INTERNAL STANDARD	VERIFIED	00:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:06
BLANK TEST	0.00	00:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:07
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:08
BLANK TEST	0.00	00:09



505 Dept of Penitentiaries WCSO #162 04-01-13