

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/02/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13050

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:08
INTERNAL STANDARD	VERIFIED	01:08
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:09
BLANK TEST	0.00	01:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:10
BLANK TEST	0.00	01:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:11
BLANK TEST	0.00	01:12



565 RJK J. Bartlett WCSO #162 0502-13