

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/08/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13050

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:29
INTERNAL STANDARD	VERIFIED	06:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:29
BLANK TEST	0.00	06:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:30
BLANK TEST	0.00	06:31
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	06:31
BLANK TEST	0.00	06:32



SGT *Ralph Bartlett* WLSU #162 05-08-13