

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 05/21/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13050

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 01:10 |
| INTERNAL STANDARD       | VERIFIED | 01:11 |
| REFERENCE STANDARD TEMP | 33.98c   |       |
| REFERENCE STANDARD      | 0.09     | 01:11 |
| BLANK TEST              | 0.00     | 01:12 |
| REFERENCE STANDARD TEMP | 33.94c   |       |
| REFERENCE STANDARD      | 0.09     | 01:12 |
| BLANK TEST              | 0.00     | 01:13 |
| REFERENCE STANDARD TEMP | 33.96c   |       |
| REFERENCE STANDARD      | 0.09     | 01:13 |
| BLANK TEST              | 0.00     | 01:14 |



505 *Ralph Bartlett* WLSO #166 05-21-13