

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 05/27/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:18
INTERNAL STANDARD	VERIFIED	01:18
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:20
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	01:21
BLANK TEST	0.00	01:21



SGS RJA/Bartlett WCO #102 05-27-13