

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/24/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:26
INTERNAL STANDARD	VERIFIED	02:26
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:26
BLANK TEST	0.00	02:27
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:27
BLANK TEST	0.00	02:28
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:28
BLANK TEST	0.00	02:29



SG- *Ralph Bartlett* WSO #162 06-24-13