

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/10/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:34
INTERNAL STANDARD	VERIFIED	00:34
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:34
BLANK TEST	0.00	00:35
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:35
BLANK TEST	0.00	00:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:36
BLANK TEST	0.00	00:37



SGS *RJP* of *Bartlett*

WESO #10c

06-10-13