

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/18/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:06
INTERNAL STANDARD	VERIFIED	00:06
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:07
BLANK TEST	0.00	00:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:08
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:09
BLANK TEST	0.00	00:09



SOS *Ralph Bartlett* WJSC #162 06-18-13