

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/08/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:44
INTERNAL STANDARD	VERIFIED	01:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:45
BLANK TEST	0.00	01:45
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	01:46
BLANK TEST	0.00	01:46
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:47



SG Ref. of *R Bartlett* WCSO #162 07-08-13