

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/17/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:21
INTERNAL STANDARD	VERIFIED	01:22
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:22
BLANK TEST	0.00	01:23
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:23
BLANK TEST	0.00	01:24
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:24
BLANK TEST	0.00	01:25



SGT Sgt J Bartlett weso #162 07-17-13