

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/24/2013

Operator Name: JOSHUA M LOPEZ
Permit #: 34728
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13090

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:24
INTERNAL STANDARD	VERIFIED	20:24
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	20:24
BLANK TEST	0.00	20:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:25
BLANK TEST	0.00	20:26
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	20:26
BLANK TEST	0.00	20:27

J.C. Jm L #299

