

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/30/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:01
INTERNAL STANDARD	VERIFIED	04:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:01
BLANK TEST	0.00	04:02
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	04:02
BLANK TEST	0.00	04:03
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	04:03
BLANK TEST	0.00	04:04



SUS Ralph J Bartlett

WCSO #102

07-30-13