

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/13/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:04
INTERNAL STANDARD	VERIFIED	01:05
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:05
BLANK TEST	0.00	01:06
REFERENCE STANDARD TEMP	33.93c	
REFERENCE STANDARD	0.09	01:06
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:07
BLANK TEST	0.00	01:08



SG Bartlett

WCSO #162

08-13-13