

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 08/19/2013

Operator Name: CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:53
INTERNAL STANDARD	VERIFIED	18:53
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	18:54
BLANK TEST	0.00	18:54
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	18:55
BLANK TEST	0.00	18:56
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	18:56
BLANK TEST	0.00	18:57

*Charles B. Geisler #248*

