

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/27/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:51
INTERNAL STANDARD	VERIFIED	00:51
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:52
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:53
BLANK TEST	0.00	00:54
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:54
BLANK TEST	0.00	00:55



SGS R.J. Bartlett

WCSO #162 08-27-13