

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 09/02/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:31
INTERNAL STANDARD	VERIFIED	03:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:31
BLANK TEST	0.00	03:32
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:32
BLANK TEST	0.00	03:33
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	03:33
BLANK TEST	0.00	03:34



SAF *Ralph Bartlett* 9-2-13

WCSO #162