

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/10/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:46
INTERNAL STANDARD	VERIFIED	01:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:48
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:48
BLANK TEST	0.00	01:49
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:49
BLANK TEST	0.00	01:50



SLI R/L J Bartlett Waco #160

09-10-13