

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 09/16/2013

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:14
INTERNAL STANDARD	VERIFIED	02:14
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:14
BLANK TEST	0.00	02:15
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:15
BLANK TEST	0.00	02:16
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:16
BLANK TEST	0.00	02:17



SGT *Ralph Bartlett* WCSO #162 09-16-13