

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/24/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13140

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:39
INTERNAL STANDARD	VERIFIED	00:39
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	00:39
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:42



SGS Rpt of Results WMSO #162 09-24-13