

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/30/2013

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 13200

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:08
INTERNAL STANDARD	VERIFIED	00:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:09
BLANK TEST	0.00	00:10
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:10
BLANK TEST	0.00	00:11
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:11
BLANK TEST	0.00	00:12



SAT *RJ Bartlett* WESCO SHERIFF #162 09/30/13